

# MAXIMA



# ESU

# TECHNICAL SPECIFICATION

MODEL	MAXIMA
Power	400 W
Operating Frequency	480 KHz
Input Voltage	170-260 VAC/50 Hz
Display	Touch Screen
Weight	5.5 kg
Program	100
Tissue Response Circuit	Available
Patient plate Monitoring System	Available
<b>Monopolar cut mode</b>	
Pure Cut	400 W
Precise cut	300 W
Blend 1	250 W
Blend 2	225 W
Blend 3	200 W
Blend 4	175 W
Blend 5	150 W
Endocut 1	150 W
Endocut 2	100 W
Endocut 3	75 W

# TECHNICAL SPECIFICATION


Monopolar coagulation modes	
Spray	120 W
Coarse Spray	100 W
Fine Spray	150 W
Fulguration	150 W
Coarse Fulguration	150 W
Fine Fulguration	150 W
Desiccation/soft	150 W
Bipolar modes	
Micro	80 W
Macro	120 W
Bi-Coag	120 W
Bi-Cut	120 W



# PERFORMANCE FEATURES

Tissue response  
circuit

Patient plate  
monitoring system



# MAXIMA

